EC Declaration of Conformity



Manufacturer:

iTs Designs Ltd Atherstone House Merry Lees Industrial Estate Desford Leicestershire LE3 3LR England

Issue Date	29/9/2008	Updated 20/12/16
Device(s)	P135A	Short Range Personal Alarm Badge
	P135AB	Long Range Personal Alarm Badge
	P135E	Panic call unit
Accessories	P160B	Li Battery, Size CR3025
	P117C	Radio linked Autodialler
	P137	Pager
	P168	Pager
Description	The P135 is a small monitor badge worn by the user. Depending on configuration	
	programming it can be used to detect panic, falling, hyperthermia risk, wandering. The	
	alarms are transmitted by Safelink Radio to an Alert-iT Pager	
Directives	2014/53/EU	RTT&E Directive
	2011/65/EU	Permitted Materials
Classification		
Standards	EN 14971:2012	Risk Assessment
Applied	EN 60950-1:2006+A2:2013	Safety
	EN 60601-1-2:2015	Medical Device EMC
	EN55011:2016	ISM Radio Disturbance
	EN12182:2012	Assistive Technology
	EN 300 220-1 V2.4.1(2012-05)	Permitted radio transmission
	EN60601-1-8:2007+A1:2013	Medical Device Alarms
	ISO22196:2011	Anti-microbial treatment

Declaration:

We certify that the apparatus to which this declaration relates conforms with the relevant requirements set out in the above referenced EC Directive(s) and standards

Signatory:

D Godfrey BSc(Hons) FIEE Ceng Managing Director