

# EC Declaration of Conformity



**Manufacturer:**

iTs Designs Ltd  
 Atherstone House  
 Merry Lees Industrial Estate  
 Desford  
 Leicestershire LE3 3LR  
 England

Issue Date	14/9/2011	updated 19/12/16
Device(s)	P166B	Wired Guide Light Monitor
Accessories	P143C P143G P144C P171 P159B P145 Miscellaneous switch devices	Bed Mat Occupancy Sensor Bed Ribbon Occupancy Sensor Bed Leg Occupancy Sensor Range of power supplies Bed/Wall Mounting Bracket Range of Nurse Call Leads Passive switching devices covered by EN12182
Description	The P166 models cover a range of small monitors that detect a user leaving a bed and provides a reassuring guide light for safety. An alarm can be transmitted either instantly or after a programmed. The alarms are transmitted by Safelink Radio (version A) or to a connected Nurse Call or Telecare system (version B)	
Directives	2014/30/EU 2014/35/EU 2011/65/EU	EMC Directive Low Voltage Permitted Materials (UK)
Classification	Safety Class II	
Standards Applied	EN 14971:2012 EN 60950-1:2006+A2:2013 EN 60601-1-2:2015 EN 12182:2012	Risk Assessment Safety Medical Device EMC Assistive Technology

**Declaration:**

We certify that the apparatus to which this declaration relates conforms with the relevant requirements set out in the above referenced EC Directive(s) and standards

**Signatory:**

D Godfrey BSc(Hons) FIEE Ceng  
 Managing Director