

# EC Declaration of Conformity



**Manufacturer:**  
iTs Designs Ltd  
Atherstone House  
Merry Lees Industrial Estate  
Desford  
Leicestershire LE3 3LR  
England

Issue Date	07/05/2014	20/12/16
Device(s)	P177B	Wired Motion Monitor
Accessories	P137/8 P117B P145	Pager Wired Autodialler Range of Nurse Call Leads
Description	The P177 detects the motion of a magnet placed on the users person or clothing. If motion is detected the alarms are transmitted by Safelink Radio or to a connected Nurse Call or Telecare system	
Directives	2014/30/EU 2014/35/EU 2011/65/EU	EMC Directive Low Voltage Permitted Materials
Classification		
Standards Applied	EN 14971:2012 EN 60950-1:2006+A2:2013 EN 60601-1-2:2015 EN 12182:2012	Risk Assessment Safety Medical Device EMC Assistive Technology

**Declaration:**

We certify that the apparatus to which this declaration relates conforms with the relevant requirements set out in the above referenced EC Directive(s) and standards

**Signatory:**

A handwritten signature in black ink, appearing to read "D Godfrey". The signature is stylized and includes a long horizontal stroke at the end.

D Godfrey BSc(Hons) FIEE Ceng  
Managing Director