

EC Declaration of Conformity



Manufacturer:

iTs Designs Ltd
 Atherstone House
 Merry Lees Industrial Estate
 Desford
 Leicestershire LE3 3LR
 England

Issue Date	20/6/2011	Updated 22/1/17
Device(s)	P155B	Radio Linked Nurse Call Interface
Accessories	P171	Range of power supplies for P155
Description	The P155 is a receiver unit for alarms transmitted using the Alert-iT Safelink Protocol. On receipt of an alarm the internal relay is operated which, depending on the connecting lead used, can trigger many different Nurse Call and Telecare systems	
Directives	2011/65/EU 2014/53/EU	Permitted Materials RTT&E Directive
Classification		
Standards Applied	EN 14971:2012 EN 60950-1:2006+A2:2013 EN 60601-1-2:2015 EN 55011:2016 EN 12182:2012 EN 300 220-1 V2.4.1(2012-05)	Risk Assessment Safety Medical Device EMC ISM Radio Disturbance Assistive Technology Radio Transmissions: Short Range Devices Supply

Declaration:

We certify that the apparatus to which this declaration relates conforms with the relevant requirements set out in the above referenced EC Directive(s) and standards

Signatory:

D Godfrey BSc(Hons) FIEE Ceng
 Managing Director